

III. Thermodynamics

Parameter/Range	Range	CMC ² (±)	Comments
Temperature-Hygrometer Digital temperature-Hygrometer	(10 to 50) °C (10 to 95) %RH	0.52 °C 2.3 % RH	Working standard: Temperature-Hygrometer ROTRONIC/HP22-A
Temperature & Humidity Recorder (Display Mode) Digital Temperature & Humidity Recorder (Display Mode)	(20 to 30) °C (30 to 85) % RH	0.5 °C 3.8 % RH	Working standard: Temperature-Hygrometer ESCORT/EI-HS-D-32-L

¹ This laboratory offers commercial calibration service.